

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 715 005 BL

**Graduated cylinders, 50:1 ml, blue grad., A, conf. cert.**




**Batch:** 23-22

**Serial:**

**Measuring Nr.** 56555

**Batch size:** 280

Volume:	10,000 ml	30,000 ml	50,000 ml
<b>Tolerance:</b>	± 0,5000 ml	± 0,5000 ml	± 0,5000 ml
<b>Average value:</b>	10,0814 ml	30,0149 ml	49,9587 ml
<b>Deviation:</b>	0,0570 ml	0,0480 ml	0,0453 ml
<b>Precision (CV):</b>	0,5705 %	0,1600 %	0,0907 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	4
<b>Date:</b>	14.06.2022
<b>Balance:</b>	Sartorius - MSE224S-0CE -DU

