

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 715 010 USP

**Graduated cylinders, 100:1 ml, USP, A, conf. cert.**




**Batch:** 29-23

**Serial:**

**Measuring Nr.:** 63831

**Batch size:** 100

Volume:	20,000 ml	60,000 ml	100,000 ml
<b>Tolerance:</b>	± 0,5000 ml	± 0,5000 ml	± 0,5000 ml
<b>Average value:</b>	20,0267 ml	60,0392 ml	99,9979 ml
<b>Deviation:</b>	0,0654 ml	0,0679 ml	0,0533 ml
<b>Precision (CV):</b>	0,3272 %	0,1131 %	0,0533 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	7
<b>Date:</b>	21.07.2023
<b>Balance:</b>	Sartorius - LE244S-OCE

