

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 715 010

Graduated cylinder, 100:1,0 ml, A, conf. cert.




Batch: 09-23

Serial:

Measuring Nr. 61919

Batch size: 1137

Volume:	20,000 ml	60,000 ml	100,000 ml
Tolerance:	± 0,5000 ml	± 0,5000 ml	± 0,5000 ml
Average value:	19,9721 ml	59,9261 ml	99,9601 ml
Deviation:	0,1129 ml	0,0959 ml	0,0576 ml
Precision (CV):	0,5644 %	0,1598 %	0,0576 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	10.03.2023
Balance:	Sartorius - LE244S-OCE

