

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 715 100

Graduated cylinder, 1000:10 ml, A, conf. cert.




Batch: 41-23

Serial:

Measuring Nr.: 64824

Batch size: 350

Volume:	200,000 ml	600,000 ml	1000,000 ml
Tolerance:	± 5,0000 ml	± 5,0000 ml	± 5,0000 ml
Average value:	200,0697 ml	600,1243 ml	1000,2034 ml
Deviation:	0,4700 ml	0,4120 ml	0,4252 ml
Precision (CV):	0,2350 %	0,0687 %	0,0425 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	26.10.2023
Balance:	Sartorius - MSE14202S- OCE-D0

