

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 715 100

Graduated cylinder, 1000:10 ml, A, conf. cert.




Batch: 24-26

Serial:

Measuring Nr. 78474

Batch size: 400

Volume:	200,0000 ml	600,0000 ml	1000,0000 ml
Tolerance:	$\pm 5,0000$ ml	$\pm 5,0000$ ml	$\pm 5,0000$ ml
Average value:	200,4172 ml	600,4208 ml	999,9092 ml
Deviation:	0,6099 ml	0,6614 ml	0,6537 ml
Precision (CV):	0,3049 %	0,1102 %	0,0654 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

4

Date:

16.06.2026

Balance:

Sartorius -
MSE14202S-
OCE-D0

DE-M