

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 715 200

Graduated cylinder, 2000:20 ml, A, conf. cert.




Batch: 17-23

Serial:

Measuring Nr.: 62805

Batch size: 105

Volume:	400,000 ml	1200,000 ml	2000,000 ml
Tolerance:	± 10,0000 ml	± 10,0000 ml	± 10,0000 ml
Average value:	399,1416 ml	1200,2150 ml	2001,4515 ml
Deviation:	0,5367 ml	0,8725 ml	0,5115 ml
Precision (CV):	0,1342 %	0,0727 %	0,0256 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	27.04.2023
Balance:	Sartorius - LE244S-OCE

