

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 754 010

Grad. cylinders, 100:1 ml, NS 24, A, conf. cert.




Batch: 47-23

Serial:

Measuring Nr. 65438

Batch size: 260

Volume:	20,000 ml	60,000 ml	100,000 ml
Tolerance:	± 0,5000 ml	± 0,5000 ml	± 0,5000 ml
Average value:	19,9148 ml	59,9744 ml	99,9763 ml
Deviation:	0,0664 ml	0,0555 ml	0,0513 ml
Precision (CV):	0,3321 %	0,0925 %	0,0513 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	06.12.2023
Balance:	Sartorius - LE244S-OCE

