

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 642 002

Volumetric flask, 20 ml, GL 14, A, conf. cert.

Batch: 11-19


Serial:

Measuring Nr. 29166

Batch size:

99

Volume:	20,000 ml
Tolerance:	$\pm 0,0400$ ml
Average value:	19,9985 ml
Deviation:	0,0051 ml
Precision (CV):	0,0256 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

0

Date:

08.04.2019

Balance:

Sartorius -
LE244S-OCE

DE-M