

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 642 100

Volumetric flask, 1000 ml, GL 32, A, conf. cert.

Batch: 44-25


Serial:

Measuring Nr. 75142

Batch size:

15

Volume:	1000,0000 ml
Tolerance:	$\pm 0,4000$ ml
Average value:	1000,0926 ml
Deviation:	0,0209 ml
Precision (CV):	0,0021 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

29.10.2025

Balance:

Sartorius -
MSE14202S-
OCE-D0

DE-M