

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 662 001

**Volumetric flask, 10 ml, NS 7, A, conf. cert.**


**Batch:** 09-23

**Serial:**

**Measuring Nr.** 61877

**Batch size:** 1792

<b>Volume:</b>	<b>10,0000 ml</b>
<b>Tolerance:</b>	$\pm 0,0250$ ml
<b>Average value:</b>	10,0003 ml
<b>Deviation:</b>	0,0029 ml
<b>Precision (CV):</b>	0,0293 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:** 7

**Date:** 06.03.2023

**Balance:** Sartorius -  
LE244S-OCE

DE-M