

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 00114 BL

Vol. flask, 10 ml, NS 14, blue grad., A, conf. cert.


Batch: 12-22

Serial:

Measuring Nr. 54564

Batch size: 500

Volume:	10,000 ml
Tolerance:	$\pm 0,0600$ ml
Average value:	9,9978 ml
Deviation:	0,0031 ml
Precision (CV):	0,0309 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	26.04.2022
Balance:	Sartorius - LE244S-OCE

DE-M