

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 662 00114

**Volumetric flasks, 10 ml, NS 14, A, conf. cert.**

**Batch:** 07-22


**Serial:**

**Measuring Nr.** 52215

**Batch size:**

500

|                        |                   |
|------------------------|-------------------|
| <b>Volume:</b>         | <b>10,0000 ml</b> |
| <b>Tolerance:</b>      | $\pm 0,0600$ ml   |
| <b>Average value:</b>  | 10,0015 ml        |
| <b>Deviation:</b>      | 0,0031 ml         |
| <b>Precision (CV):</b> | 0,0313 %          |



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:**

7

**Date:**

22.02.2022

**Balance:**

Sartorius -  
LE244S-OCE

DE-M