

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 002

Volumetric flasks, 20 ml, NS 10, A, conf. cert.

Batch: 33-19


Serial:

Measuring Nr. 32855

Batch size:

2496

Volume:	20,000 ml
Tolerance:	$\pm 0,0400$ ml
Average value:	20,0032 ml
Deviation:	0,0043 ml
Precision (CV):	0,0215 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

09.09.2019

Balance:

Sartorius -
LE244S-OCE

DE-M