

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 0075

Volumetric flask, 75 ml, NS 12, A, conf. cert.


Batch: 47-25

Serial:

Measuring Nr. 75383

Batch size: 150

Volume:	75,0000 ml
Tolerance:	$\pm 0,0600$ ml
Average value:	75,0262 ml
Deviation:	0,0072 ml
Precision (CV):	0,0096 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	18.11.2025
Balance:	Sartorius - MSE224S-0CE -DU

DE-M