

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 010 Bates

Volumetric flasks, 100 ml, crimped edge, A, conf. cert.

Batch: 51-18


Serial:

Measuring Nr. 26643

Batch size:

280

Volume:	100,000 ml
Tolerance:	$\pm 0,1000$ ml
Average value:	99,9857 ml
Deviation:	0,0167 ml
Precision (CV):	0,0167 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

9

Date:

20.12.2018

Balance:

Sartorius -
LE244S-OCE

DE-M