

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 662 010

**Vol. flasks, 100 ml, NS 12, A, conf. cert.**

**Batch:** 44-17


**Serial:** -

**Measuring Nr.** 15799

**Batch size:**

3200

<b>Volume:</b>	<b>100,000 ml</b>
<b>Tolerance:</b>	$\pm 0,1000$ ml
<b>Average value:</b>	100,0009 ml
<b>Deviation:</b>	0,0160 ml
<b>Precision (CV):</b>	0,0160 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:**

4

**Date:**

20.11.2017

**Balance:**

Sartorius -  
LE244S-OCE

DE-M