

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 010

Vol. flasks, 100 ml, NS 12, A, conf. cert.

Batch: 44-17


Serial:

Measuring Nr. 17636

Batch size:

1

Volume:	100,000 ml
Tolerance:	± 0,1000 ml
Measured value:	99,9988 ml



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

9

Date:

29.01.2018

Balance:

Sartorius -
LE244S-OCE

DE-M