

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 662 011 BL

**Vol. flasks, 100 ml, NS 14, blue grad., A, conf. cert.**


**Batch:** 14-20

**Serial:**

**Measuring Nr.** 37190

**Batch size:** 965

<b>Volume:</b>	<b>100,000 ml</b>
<b>Tolerance:</b>	$\pm 0,1000$ ml
<b>Average value:</b>	100,0120 ml
<b>Deviation:</b>	0,0140 ml
<b>Precision (CV):</b>	0,0140 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	4
<b>Date:</b>	03.04.2020
<b>Balance:</b>	Sartorius - MSE224S-0CE -DU

DE-M