

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 012 BL

Vol. flasks, 10 ml, NS 10, blue grad., A, conf. cert.


Batch: 17-26

Serial:

Measuring Nr. 77601

Batch size: 230

Volume:	10,0000 ml
Tolerance:	$\pm 0,0400$ ml
Average value:	9,9927 ml
Deviation:	0,0042 ml
Precision (CV):	0,0423 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator: 7

Date: 21.04.2026

Balance: Sartorius -
LE244S-OCE

DE-M