Certificate of performance


witeg GERMANY

ISO 9001 CERTIFIED

Article:	**3 662 015 BL**
Volumetric flasks, 5 ml, NS 10, blue grad., A, conf. cert.	

Batch:	38-26	**Serial:**	
Measuring Nr.	80382	**Batch size:**	499

Volume:	**5,0000 ml**
Tolerance:	± 0,0400 ml
Average value:	5,0026 ml
Deviation:	0,0032 ml
Precision (CV):	0,0634 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH
Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	22.10.2026
Balance:	Sartorius - LE244S-0CE

DE-M