

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 016

Volumetric flasks, 15 ml, NS 10, A, conf. cert.


Batch: 43-18

Serial: -

Measuring Nr. 27735

Batch size: 1

Volume:	15,0000 ml
Tolerance:	$\pm 0,0400$ ml
Measured value:	14,9912 ml



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	9
Date:	15.02.2019
Balance:	Sartorius - LE244S-OCE

DE-M