

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 020 KU

Volumetric flask, 200 ml, NS 14, KU, A, conf. cert.


Batch: 1

Serial:

Measuring Nr. 47790

Batch size: 20

Volume:	200,000 ml
Tolerance:	$\pm 0,1500$ ml
Average value:	200,0206 ml
Deviation:	0,0027 ml
Precision (CV):	0,0013 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	08.07.2021
Balance:	Sartorius - LA 1200 S

DE-M