

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 025 BL

Vol. flasks, 250 ml, NS 14, blue grad., A, conf. cert.


Batch: 41-17

Serial:

Measuring Nr. 14909

Batch size: 2400

Volume:	250,0000 ml
Tolerance:	$\pm 0,1500$ ml
Average value:	250,0096 ml
Deviation:	0,0208 ml
Precision (CV):	0,0083 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	16.10.2017
Balance:	Sartorius - LE244S-OCE

DE-M