

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 662 025 BL

**Vol. flask, 250 ml, NS 14, blue grad., A, conf. cert.**


**Batch:** 09-22

**Serial:**

**Measuring Nr.** 53011

**Batch size:** 490

<b>Volume:</b>	250,000 ml
<b>Tolerance:</b>	$\pm 0,1500$ ml
<b>Average value:</b>	250,0015 ml
<b>Deviation:</b>	0,0120 ml
<b>Precision (CV):</b>	0,0048 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:** 4

**Date:** 14.03.2022

**Balance:** Sartorius - LA  
1200 S

DE-M