

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 025

Volumetric flasks, 250 ml, NS 14, A, conf. cert.

Batch: 42-17


Serial:

Measuring Nr. 15448

Batch size:

1200

Volume:	250,0000 ml
Tolerance:	$\pm 0,1500$ ml
Average value:	249,9849 ml
Deviation:	0,0240 ml
Precision (CV):	0,0096 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

4

Date:

26.10.2017

Balance:

Sartorius - LA
1200 S

DE-M