

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 662 050

**Volumetric flask, 500 ml, NS 19, A, conf. cert.**

**Batch:** 16-26


**Serial:**

**Measuring Nr.** 77478

**Batch size:**

697

<b>Volume:</b>	500,000 ml
<b>Tolerance:</b>	$\pm 0,2500$ ml
<b>Average value:</b>	500,0169 ml
<b>Deviation:</b>	0,0369 ml
<b>Precision (CV):</b>	0,0074 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:**

4

**Date:**

16.04.2026

**Balance:**

Sartorius - LA  
1200 S

DE-M