

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 090 BL

Vol. flask., 900 ml, NS 24, blue grad., A, conf. cert.


Batch: 32-25

Serial:

Measuring Nr. 74093

Batch size: 100

Volume:	900,000 ml
Tolerance:	$\pm 0,4000$ ml
Average value:	900,0541 ml
Deviation:	0,0324 ml
Precision (CV):	0,0036 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	05.08.2025
Balance:	Sartorius - LA 1200 S

DE-M