

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 100 BL USP

Vol. flask, 1000 ml, NS 24, blue, USP, A, conf. cert.

Batch: 45-18


Serial:

Measuring Nr. 25682

Batch size:

50

Volume:	1000,0000 ml
Tolerance:	$\pm 0,3000$ ml
Average value:	999,9272 ml
Deviation:	0,0760 ml
Precision (CV):	0,0076 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

4

Date:

20.11.2018

Balance:

Sartorius - LC
6200 S

DE-M