

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 200 BL

Vol. flask, 2000 ml, NS 29, blue grad., A, conf. cert.


Batch: 43-23

Serial:

Measuring Nr. 64955

Batch size: 136

Volume:	2000,0000 ml
Tolerance:	$\pm 0,6000$ ml
Average value:	2000,1976 ml
Deviation:	0,0490 ml
Precision (CV):	0,0024 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	10.11.2023
Balance:	Sartorius - MSE14202S- OCE-D0

DE-M