

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 200

Volumetric flask, 2000 ml, NS 29, A, conf. cert.


Batch: 35-24

Serial:

Measuring Nr. 68946

Batch size: 490

Volume:	2000,0000 ml
Tolerance:	$\pm 0,6000$ ml
Average value:	2000,0664 ml
Deviation:	0,0616 ml
Precision (CV):	0,0031 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	06.09.2024
Balance:	Sartorius - MSE14202S- OCE-D0

DE-M