

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 300 BL

Vol. flask, 3000 ml, NS 29, blue grad., A, conf. cert.


Batch: 05-19

Serial:

Measuring Nr. 26949

Batch size: 15

Volume:	3000,0000 ml
Tolerance:	± 0,8000 ml
Average value:	3000,3752 ml
Deviation:	0,1209 ml
Precision (CV):	0,0040 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	23.01.2019
Balance:	Sartorius - MSE14202S- OCE-D0

DE-M