

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 500 BL

Vol. flask, 5000 ml, NS 34, blue grad., A, conf. cert.

Batch: 09-24


Serial:

Measuring Nr. 66310

Batch size:

37

Volume:	5000,0000 ml
Tolerance:	$\pm 1,2000$ ml
Average value:	4999,9706 ml
Deviation:	0,3756 ml
Precision (CV):	0,0075 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

4

Date:

11.03.2024

Balance:

Sartorius -
MSE14202S-
OCE-D0

DE-M