

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 600

Volumetric flask, 6000 ml, NS 34, A, conf.cert.

Batch: 42-25


Serial:

Measuring Nr. 75198

Batch size:

10

Volume:	6000,0000 ml
Tolerance:	$\pm 1,2000$ ml
Average value:	6000,6527 ml
Deviation:	0,0423 ml
Precision (CV):	0,0007 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

4

Date:

05.11.2025

Balance:

Sartorius -
MSE14202S-
OCE-D0

DE-M