

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 669 001 BR

**Vol. flasks, 2 ml, NS 7, amber glass, A, conf. cert.**


**Batch:** 02-23

**Serial:**

**Measuring Nr.** 61240

**Batch size:** 99

<b>Volume:</b>	2,0000 ml
<b>Tolerance:</b>	$\pm 0,0250$ ml
<b>Average value:</b>	2,0011 ml
<b>Deviation:</b>	0,0021 ml
<b>Precision (CV):</b>	0,1075 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:** 7

**Date:** 25.01.2023

**Balance:** Sartorius -  
LE244S-OCE

DE-M