

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 669 00210 BL

Vol. flask, 3 ml, NS 10, trapezoidal, A, conf.cert.

Batch: 13-26


Serial:

Measuring Nr. 77166

Batch size:

280

Volume:	3,0000 ml
Tolerance:	$\pm 0,0400$ ml
Average value:	2,9988 ml
Deviation:	0,0025 ml
Precision (CV):	0,0830 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

30.03.2026

Balance:

Sartorius -
LE244S-OCE

DE-M