

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 670 001

Vol. flask, 10 ml, NS 7, amber glass, A, conf. cert.

Batch: 32-19


Serial:

Measuring Nr. 32336

Batch size:

1185

| | |
|------------------------|------------------|
| Volume: | 10,000 ml |
| Tolerance: | $\pm 0,0250$ ml |
| Average value: | 10,0047 ml |
| Deviation: | 0,0046 ml |
| Precision (CV): | 0,0463 % |



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

4

Date:

27.08.2019

Balance:

Sartorius -
LE244S-OCE

DE-M