

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 670 006

Vol. flasks, 50 ml, NS 14, amber glass, A, conf. cert.


Batch: 23-20

Serial:

Measuring Nr. 38736

Batch size: 500

Volume:	50,000 ml
Tolerance:	$\pm 0,1000$ ml
Average value:	50,0023 ml
Deviation:	0,0099 ml
Precision (CV):	0,0199 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	16.06.2020
Balance:	Sartorius - MSE224S-0CE -DU

DE-M