

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 670 010

Vol. flask, 100 ml, NS 12, amber glass, A, conf. cert.

Batch: 04-25


Serial:

Measuring Nr. 71873

Batch size:

1199

Volume:	100,000 ml
Tolerance:	$\pm 0,1000$ ml
Average value:	99,9577 ml
Deviation:	0,0203 ml
Precision (CV):	0,0203 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

30.01.2025

Balance:

Sartorius -
LE244S-OCE

DE-M