

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 670 012 USP

Vol. flasks, 10 ml, NS 10, amber glass, A, conf. cert.

Batch: 51-25


Serial:

Measuring Nr. 76619

Batch size:

327

Volume:	10,0000 ml
Tolerance:	± 0,0200 ml
Average value:	10,0070 ml
Deviation:	0,0049 ml
Precision (CV):	0,0487 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

4

Date:

11.02.2026

Balance:

Sartorius -
MSE224S-0CE
-DU

DE-M