

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 670 015

Vol. flasks, 5 ml, NS 10, amber glass, A, conf. cert.


Batch: 22-22

Serial:

Measuring Nr. 56912

Batch size: 280

Volume:	5,0000 ml
Tolerance:	$\pm 0,0400$ ml
Average value:	4,9928 ml
Deviation:	0,0040 ml
Precision (CV):	0,0797 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator: 7

Date: 30.06.2022

Balance: Sartorius -
LE244S-OCE

DE-M