

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 670 100

Vol. flask, 1000 ml, NS 24, amber glass, A, conf. cert.

Batch: 12-23


Serial:

Measuring Nr. 62035

Batch size:

254

Volume:	1000,0000 ml
Tolerance:	± 0,4000 ml
Average value:	1000,0684 ml
Deviation:	0,0906 ml
Precision (CV):	0,0091 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

4

Date:

21.03.2023

Balance:

Sartorius -
MSE14202S-
OCE-D0

DE-M