

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 510 051

**Graduated pipettes, 0,5:0,01 ml, AS, conf. cert.**

**Batch:** 49-22




**Serial:**

**Measuring Nr.** 60923

**Batch size:**

500

Volume:	0,5000 ml	0,3000 ml	0,1000 ml
<b>Tolerance:</b>	± 0,0060 ml	± 0,0060 ml	± 0,0060 ml
<b>Average value:</b>	0,4994 ml	0,3012 ml	0,1001 ml
<b>Deviation:</b>	0,0017 ml	0,0013 ml	0,0007 ml
<b>Precision (CV):</b>	0,3484 %	0,4253 %	0,6951 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 835

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:**

7

**Date:**

11.01.2023

**Balance:**

Sartorius -  
LE244S-OCE

DE-M