

Certificate of performance



Article: 3 510 101

Grad. pipettes, 1:0,01 ml, AS, conf. cert.




Batch: 14-23

Serial:

Measuring Nr. 62622

Batch size: 3200

Volume:	1,0000 ml	0,6000 ml	0,2000 ml
Tolerance:	± 0,0070 ml	± 0,0070 ml	± 0,0070 ml
Average value:	0,9987 ml	0,5997 ml	0,2002 ml
Deviation:	0,0018 ml	0,0016 ml	0,0011 ml
Precision (CV):	0,1772 %	0,2660 %	0,5652 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 835

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	21.04.2023
Balance:	Sartorius - LE244S-OCE

