

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 510 202

Grad. pipettes, 2:0,02 ml, AS, conf. cert.




Batch: 12-19

Serial:

Measuring Nr. 28736

Batch size: 2400

Volume:	2,0000 ml	1,2000 ml	0,4000 ml
Tolerance:	± 0,0100 ml	± 0,0100 ml	± 0,0100 ml
Average value:	1,9990 ml	1,2036 ml	0,4016 ml
Deviation:	0,0026 ml	0,0026 ml	0,0020 ml
Precision (CV):	0,1284 %	0,2140 %	0,5002 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 835

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	0
Date:	28.03.2019
Balance:	Sartorius - LE244S-OCE

