

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 510 505

Graduated pipettes, 5:0,05 ml, AS, conf. cert.




Batch: 08-22

Serial:

Measuring Nr. 52614

Batch size: 3199

Volume:	5,0000 ml	3,0000 ml	1,0000 ml
Tolerance:	± 0,0300 ml	± 0,0300 ml	± 0,0300 ml
Average value:	4,9994 ml	2,9911 ml	0,9994 ml
Deviation:	0,0073 ml	0,0043 ml	0,0023 ml
Precision (CV):	0,1450 %	0,1436 %	0,2309 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 835

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	01.03.2022
Balance:	Sartorius - LE244S-OCE

