

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 510 710

**Graduated pipettes, 10:0,1 ml, AS, conf. cert.**




**Batch:** 08-21

**Serial:**

**Measuring Nr.** 44802

**Batch size:** 5000

Volume:	10,000 ml	6,0000 ml	2,0000 ml
<b>Tolerance:</b>	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
<b>Average value:</b>	9,9879 ml	5,9991 ml	1,9964 ml
<b>Deviation:</b>	0,0069 ml	0,0093 ml	0,0063 ml
<b>Precision (CV):</b>	0,0692 %	0,1548 %	0,3149 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 835

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	7
<b>Date:</b>	04.03.2021
<b>Balance:</b>	Sartorius - LE244S-OCE

