

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 515 810

Grad. pipettes, 20:0,1 ml, AS, conf. cert.




Batch: 44-20

Serial:

Measuring Nr. 42208

Batch size: 138

Volume:	20,000 ml	12,000 ml	4,000 ml
Tolerance:	± 0,1000 ml	± 0,1000 ml	± 0,1000 ml
Average value:	19,9769 ml	11,9872 ml	4,0002 ml
Deviation:	0,0218 ml	0,0114 ml	0,0098 ml
Precision (CV):	0,1089 %	0,0954 %	0,2456 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 835

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	04.11.2020
Balance:	Sartorius - LE244S-OCE

