

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 535 010 BL

**Volumetric pipettes, 1 ml, blue grad., AS, conf. cert.**


**Batch:** 32-25

**Serial:**

**Measuring Nr.** 74913

**Batch size:** 474

<b>Volume:</b>	1,0000 ml
<b>Tolerance:</b>	$\pm 0,0080$ ml
<b>Average value:</b>	1,0011 ml
<b>Deviation:</b>	0,0026 ml
<b>Precision (CV):</b>	0,2569 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	7
<b>Date:</b>	08.10.2025
<b>Balance:</b>	Sartorius - LE244S-OCE

DE-M