

# Certificate of performance

**Article:** 3 535 025

**Volumetric pipettes, 2,5 ml, AS, conf. cert.**

**Batch:** 06-26


**Serial:**

**Measuring Nr.** 76839

**Batch size:**

500

<b>Volume:</b>	<b>2,5000 ml</b>
<b>Tolerance:</b>	$\pm 0,0100$ ml
<b>Average value:</b>	2,5030 ml
<b>Deviation:</b>	0,0023 ml
<b>Precision (CV):</b>	0,0905 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:**

4

**Date:**

03.03.2026

**Balance:**

Sartorius -  
MSE224S-0CE  
-DU