

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 030 BL

Volumetric pipettes, 3 ml, blue grad., AS, conf. cert.


Batch: 06-26

Serial: -

Measuring Nr. 76328

Batch size: 279

Volume:	3,0000 ml
Tolerance:	$\pm 0,0100$ ml
Average value:	3,0031 ml
Deviation:	0,0016 ml
Precision (CV):	0,0532 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

4

Date:

28.01.2026

Balance:

Sartorius -
MSE224S-0CE
-DU

DE-M